

Gen 3

Lab Grade pH Probe

Double junction silver / silver chloride
with Vanadium III

Reads	pH
Range	0 – 14
Resolution	+/- 0.001
Accuracy	+/- 0.002
Response time	95% in 1s
Temperature range °C	-5 – 99 °C
Max pressure	100 PSI
Max depth	70m (230 ft)
Connector	Male SMA / Male BNC (Optional)
Cable length	1 meter
Internal temperature sensor	No
Time before recalibration	~1 Year
Life expectancy	~2.5 Years +




Operating principle

The Atlas Scientific Gen 3 Lab Grade pH probe incorporates our latest sensing technologies. They are designed to offer improved accuracy, stability, and responsiveness in challenging environmental conditions.

Sensor Technology


- The high surface area Ag/AgCl electrode provides a strong output signal, improving responsiveness in weak and rapidly changing pH environments.
- Vanadium III works as a hydrogen amplifier. This amplification reduces noise, leading to improved reading stabilization.
- A Tyvek™ junction connects the pH probe to the outside world. Because Tyvek™ resists water flow, the internal electrolyte is held in the probe longer leading to less frequent recalibrations.



The Vanadium III hydrogen amplifier will fade in sunlight, but not in efficacy.

Measurements


Storage Life ~5 Years
Working Life ~2.5 Years



Soaker bottle
~3.8 – 4.2 pH

107mm
(4.2")

153mm
(6")




NSF/ANSI 51 Compliant

Food Safe

Atlas Scientific LLC, hereby certifies that,

Lab Grade pH Probe
Part # ENV-40-pH

Complies with NSF/ANSI Standard 51



✓ **PVC**
NSF-51 Compliant

✓ **ABS**
NSF-51 Compliant

✓ **Glass**
NSF-51 Compliant

✓ **Polycarbonate**
NSF-51 Compliant

✓ **Polyethylene**
NSF-51 Compliant

✓ **Silicone Rubber**
NSF-51 Compliant

✓ **Titanium Nitride**
NSF-51 Compliant

Specifications

Reference electrode	Silver / silver chloride
Double junction	Yes
Max depth	70m (230 ft)
Cable length	1 meter
Weight	37 grams
Speed of response	95% in 1 second
Isopotential point	pH 7.00 (0 mV)
Dimensions	12mm x 150.6mm (0.5" x 6")
SMA connector	Male
Sterilization	Chemical only
Food Safe	Yes




DO NOT BOIL



DO NOT FREEZE

Materials



This pH probe can be **fully submerged** in fresh or salt water, up to the SMA connector **indefinitely**.


Typical applications

- Standard lab use
- Field use
- Soil
- High pH solutions (up to 14 pH)
- Samples containing heavy metals
- Hydroponics / aquaponics
- Beer, wine, alcohol, and food production

A pH electrode is a passive device that detects a current generated from hydrogen ion activity. This current (*which can be positive or negative*) is very weak and cannot be detected with a multimeter, or an analog to digital converter. This weak electrical signal can easily be disrupted and care should be taken to only use proper connectors and cables.



Result will **always** read zero.



Result will **always** read zero.

The current that is generated from the hydrogen ion activity is the reciprocal of that activity and can be predicted using this equation:

$$E = E^0 + \frac{RT}{F} \ln(\alpha_{H^+}) = E^0 - \frac{2.303RT}{F} pH$$


Where **R** is the ideal gas constant.

T is the temperature in Kelvin.

F is the Faraday constant.



Because a pH probe is a passive device it can pick up voltages that are transmitted through the solution being measured. This will result in incorrect readings and will slowly damage the pH probe over time. In this instance, proper isolation is required.

**NEVER EXTEND THE CABLE
WITH CHEAP JUMPER WIRES!**



**DO NOT CUT THE PROBE CABLE
WITHOUT REFERRING TO **THIS DOCUMENT!****

**DO NOT MAKE YOUR OWN
UNSHIELDED CABLES!**




ONLY USE SHIELDED CABLES.

Extending the probe cable length


You can extend the cable to greater than 100 meters with no loss of signal. Atlas Scientific has tested up to 300 meters without a problem, however you run the risk of turning your pH probe into an antenna, picking up noise along the length of your cable.

If you want to extend your cable, we recommend that you use proper isolation, such as the **Basic EZO™ Inline Voltage Isolator**, or an **i2 InterLink**. Be sure to calibrate your probe with the extended cable.

Extending a probe cable can be easily done with our **SMA Extension Cables**. Simply connect the SMA end of the probe to the Extension cable, and you are all set.




If you need to water proof a SMA connection, we highly recommend using a product like **Coax-Seal** to safely cover and prevent any water damage that may occur.



Helpful operating tips

pH probes must stay wet and cannot be allowed to dry out, this is why every pH probe is shipped in a plastic soaker bottle containing pH probe storage solution. The probe should remain in the bottle until it is used. If the probe is used infrequently, the bottle and its solution should be saved and the probe stored inside.

- 1 To remove the soaker bottle from the probe, hold the soaker bottle by the cap and turn the bottle until it separates from the cap; Then slide the cap and o-ring off.




Ready to use

- 2 During shipment the air bubble in the probes stem may move into the bulb area. If bubbles are seen in the bulb area, hold the probe by its top cap and shake downward as done with a clinical thermometer.

Bubbles



- 3** Vigorously stir the probe in the sample, calibration solution, or rinse solution. This action will bring solution to the probes surface quicker and improve the speed of response.



Response time

Probe cleaning

Coating of the pH bulb can lead to erroneous readings including shortened span (slope). The type of coating will determine the cleaning technique. Soft coatings can be removed by vigorous stirring or by the use of a squirt bottle. Organic chemical, or hard coatings, should be chemically removed using a light bleach solution. If cleaning does not restore performance, reconditioning may be tried. **Do not use a brush or abrasive materials on the pH probe.**



How often do you need to recalibrate a pH probe?

Because every use case is different, there is no set schedule for recalibration.

If you are using your probe in a fish tank, a hydroponic system or any environment that has generally weak levels of acids and bases you will only need to recalibrate your probe once per year for the first two years. After that every ~six months.

If you are using the pH probe in batch chemical manufacturing, industrial process, or in a solution that is known to have strong acids and bases, then calibration should be done monthly or in extreme cases after each batch.

Probe reconditioning

When reconditioning your pH probe is required due to aging, we recommend you use the **Atlas Scientific pH probe reconditioning kit**.



1980's — Today



Despite appearances
THE KCl CREEP
is really quite harmless.

The white crystals
you may find on your electrode
are formed by potassium chloride (KCl)
from the electrode filling solution.
Rinse the KCl from the electrode
with distilled water and proceed as usual.



Dried KCl residue
from pH storage
solution

Decades later...

KCl continues to behave the same way.

If you encounter the "KCl CREEP" rinse off your probe with water,
and carry on. **Your probe is not damaged.**